



BIDS AND AWARDS COMMITTEE

SUPPLEMENTAL BID BULLETIN

For ITB No. 2021-0106-ATQAL

PROCUREMENT FOR VARIOUS LABORATORY EQUIPMENT FOR ALABANG TESTING AND QUALITY ASSURANCE

4 June 2021

In response to the issues/ concerns raised during the Pre-Bid Conference held on 01 June 2021 for the above mentioned ITB, hereunder are the amendments, clarifications and recommendations.

Refer to the comparative table given below for the previous (from) and revised (to) specifications:

A. Under Section VII Technical Specification:

SECTION	REVISION	
	FROM	TO
SECTION VI OF BID DOCS	Various schedule of requirements ranging from 45 calendar days to 60 calendar days after receipt of Notice to Proceed	Within Ninety (90) calendar days after receipt of Notice to Proceed Note: applies to Item Nos. 1 to 8
SECTION VII, ITEM NO. 1, Technical Specification Letter B. Delivery Period	Within forty-five (45) calendar days after receipt of Notice to Proceed.	Within Ninety (90) calendar days after receipt of Notice to Proceed
SECTION VII, ITEM NO. 2, Technical Specification Letter B. Delivery Period	Within forty-five (45) calendar days after receipt of Notice to Proceed.	Within Ninety (90) calendar days after receipt of Notice to Proceed
Letter D.1 under Warranty and After Sales Service	A minimum of two (2) years warranty on parts and services of Digital Dry Bath after commissioning and acceptance with provision of annual calibration at three (3) test points by ISO 17025 accredited calibration service provider, during the warranty period.	A minimum of one (1) year warranty on parts and two (2) years warranty on services of Digital Dry Bath. Both commencing after successful installation by the Supplier and its acceptance by FDA. An annual calibration at three (3) test points by an ISO 17025 accredited calibration service provider, for the two (2) year warranty period must be provided.



Letter E.1 under Documentation	Training certificate of Service Personnel from the manufacturer related to the installation, equipment maintenance and troubleshooting must be provided	Submission of authorization or any applicable document to substantiate competency of service personnel in performing necessary services for the digital dry bath (e.g. installation, equipment maintenance and troubleshooting).
SECTION VII, ITEM NO. 3, Technical Specification Letter A. PCR Cabinet	Conforms to at least any of the following: ETL, NSF, ADA or ISO-5	Conforms to clean room standards (e.g. ETL, NSF, ADA or ISO-5)
	With front tempered glass closure and side panels	With front tempered glass closure and side panels or better type of material Note: documentation must be submitted to substantiate that material of the offered item is better than the tempered glass
Letter B.2 under Power Supply	Provision of Uninterrupted Power Supply with at least 3 KVA that can provide at least 20 mins backup at full load.	Provision of Uninterrupted Power Supply with at least 3 KVA that can provide 10 - 20 mins. backup at full load.
Letter C.1 under Accessories, Spares and Consumables	Inclusive of: - at least one (1) piece of UV Lamp - at least one (1) piece of HEPA Filter	Inclusive of: - at least one (1) piece of UV Lamp - at least one (1) piece of HEPA Filter Note: - delivery of HEPA filter should be done whenever needed by the end-user. The winning bidder will be notified at least one (1) week before the required delivery schedule. - Commitment Letter on delivery whenever necessary, must be submitted
Letter D. Delivery Period	Within sixty (60) calendar days after receipt of Notice to Proceed.	Within Ninety (90) calendar days after receipt of Notice to Proceed
Letter E. under Warranty and After Sales Service	A minimum of two (2) year warranty on parts and services of PCR Cabinet after commissioning and acceptance with provision of re-certification annually by NSF certified service provider during the warranty period (please provide scope of work of the PM visits)	A minimum of one (1) year warranty on parts and two (2) years warranty on services of PCR Cabinet. Both commencing after successful installation by the Supplier and its acceptance by FDA. An annual re-certification by NSF certified service provider, for the two (2) year warranty period must be provided.

<p>Letter I.5. under Other Requirements</p>	<p>The supplier must have at least ten (10) local installations of the above unit that can be viewed and demo to the end user, if necessary. Two (2) out of ten (10) local installations must be the same model as the one being offered</p> <p>Note: Bidder to submit list of companies with installed unit. Contact person and contact details (telephone number/mobile number and e-mail address) for each listed company must also be submitted.</p>	<p>The supplier must have at least five (5) local installations of the PCR Cabinet that can be viewed and demo to the end user, if necessary. Two (2) out of five (5) local installations must be the same model as the one being offered</p> <p>Note: Bidder to submit list of companies with installed unit. Contact person and contact details (telephone number/mobile number and e-mail address) for each listed company must also be submitted.</p>
<p>SECTION VII, ITEM NO. 4, Technical Specification Letter A. Main Unit</p>	<p>Temperature range Refrigerator: At least +2 to +20°C Freezer: At least -5 to -30°C</p>	<p>Temperature range Refrigerator: At least +2°C to +8°C Freezer: At least -10°C to -26°C</p>
<p>Inclusions under Letter A. Main Unit</p>	<p>One (1) unit refrigerator thermometer One (1) unit freezer thermometer Note: both thermometers must be calibrated at three (3) test points by ISO 17025 accredited service provider</p>	<p>Inclusions for <u>each</u> unit of Pharmaceutical Refrigerator with Freezer One (1) unit refrigerator thermometer One (1) unit freezer thermometer Note: both thermometers must be calibrated at three (3) test points by ISO 17025 accredited service provider</p>
<p>Letter C. under Delivery Period</p>	<p>Within forty-five (45) calendar days after receipt of Notice to Proceed.</p>	<p>Within Ninety (90) calendar days after receipt of Notice to Proceed</p>
<p>SECTION VII, ITEM NO. 5, Technical Specification Letter C.1 under Warranty and After Sales Service</p>	<p>A minimum of one (1) year warranty on parts and two (2) years warranty on services of the pipettors after commissioning and acceptance with provision of annual calibration for each unit at three (3) test points by an ISO 17025 service provider during the warranty period.</p>	<p>A minimum of one (1) year warranty on parts and two (2) years warranty on services of the pipettors after commissioning and acceptance with provision of calibration for each unit at three (3) test points by an ISO 17025 service provider prior to acceptance and after one year of acceptance.</p>
<p>SECTION VII, ITEM NO. 6, Technical Specification Letter A. Main Unit</p>	<p>Maximum speed (rpm): $\geq 15,000$</p>	<p>Maximum speed (rpm): $\geq 13,000$</p>
<p>Letter C. Delivery Period</p>	<p>Within forty-five (45) calendar days after receipt of Notice to Proceed.</p>	<p>Within Ninety (90) calendar days after receipt of Notice to Proceed</p>

<p>SECTION VII, ITEM NO. 7, Technical Specification Letter C.1 under Accessories, Spares and Consumables</p>	<p>Inclusive of: - at least one (1) piece of UV Lamp - at least one (1) piece of HEPA Filter</p>	<p>Inclusive of: - at least one (1) piece of UV Lamp - at least one (1) piece of HEPA Filter</p> <p>Note: - delivery of HEPA filter should be done whenever needed by the end-user. The winning bidder will be notified at least one (1) week before the required delivery schedule. - Commitment Letter on delivery whenever necessary, must be submitted</p>
<p>Letter D. Delivery Period</p>	<p>Within forty-five (45) calendar days after receipt of Notice to Proceed.</p>	<p>Within Ninety (90) calendar days after receipt of Notice to Proceed</p>
<p>Letter I.5. under Other Requirements</p>	<p>The supplier must have at least ten (10) local installations of the above unit that can be viewed and demo to the end user, if necessary. Two (2) out of ten (10) local installations must be the same model as the one being offered</p> <p>Note: Bidder to submit list of companies with installed unit. Contact person and contact details (telephone number/mobile number and e-mail address) for each listed company must also be submitted.</p>	<p>The supplier must have at least five (3) local installations of the Biological Safety Cabinet that can be viewed and demo to the end user, if necessary. Two (2) out of five (3) local installations must be the same model as the one being offered</p> <p>Note: Bidder to submit list of companies with installed unit. Contact person and contact details (telephone number/mobile number and e-mail address) for each listed company must also be submitted.</p>
<p>SECTION VII, ITEM NO. 8, Technical Specification Letter D. Delivery Period</p>	<p>Within forty-five (45) calendar days after receipt of Notice to Proceed.</p>	<p>Within Ninety (90) calendar days after receipt of Notice to Proceed</p>
<p>Letter C.2 under Accessories, Spares and Consumables</p>	<p>Maximum Speed: ≤ 3400 rpm</p>	<p>Maximum Speed: ≤ 3000 rpm</p>

<p>Letter C.4</p>	<p>Provision of the following kits: - at least five (5) Virus RNA/DNA Extraction Kits - at least one (1) COVID-19 Detection Kits Note: kits must be delivered only after the RT PCR Unit has been fully commissioned. Expiration date must be at least six (6) months upon delivery.</p>	<p>Provision of the following kits: - at least five (5) Virus RNA/DNA Extraction Kits - at least one (1) COVID-19 Detection Kits Note: a. kits must be delivered only after the RT PCR Unit has been fully commissioned. Expiration date must be at least six (6) months upon delivery. b. Kits must have FDA Special Certificate, whenever applicable</p>
<p>Letter D.1</p>	<p>- Safety feature: with dual lid switch</p>	<p>- Safety feature: with dual lid switch or any equivalent Safety feature of the equipment during its operation</p>
<p>Letter E.3 under Warranty and After Sales Services</p>	<p>Provision of two (2) preventive maintenance (PM) visits annually within the warranty period. The PM must include at minimum the cleaning, testing, checking of various functions, and carrying out Quality Control (QC) checks with calibration/re-certification of the instrument, and all PM parts and consumables (PM kits).</p> <p>There should be a post PM performance check to confirm that the system meets the manufacturer performance specification.</p> <p>Note: Details of scope must be provided.</p>	<p>Provision of annual preventive maintenance (PM) visit within the warranty period. The PM must include at minimum the cleaning, testing, checking of various functions, and carrying out Quality Control (QC) checks with calibration/re-certification of the instrument, and all PM parts and consumables (PM kits).</p> <p>There should be a post PM performance check to confirm that the system meets the manufacturer performance specification.</p> <p>Note: Details of scope must be provided.</p>

<p>Letter J.5 under Other Requirements</p>	<p>The supplier must have at least ten (10) local installations of RTPCR that can be viewed and demo to the end user, if necessary. Two (2) out of ten (10) local installations must be the same model as the one being offered.</p> <p>Note: Bidder to submit list of companies with installed unit. Contact person and contact details (telephone number/mobile number and e-mail address) for each listed company must also be submitted.</p>	<p>The supplier must have either of the following:</p> <p>(a) Ten (10) installations of RTPCR, regardless of brand. Two (2) out of ten (10) installations are done locally for the same brand and model as the offered unit.</p> <p>(b) Ten (10) local installations of RTPCR, with proof of at least two (2) installations of the same brand and model as the offered unit, either globally or locally.</p> <p>Note: Bidder to submit list of companies with installed unit. Contact person and contact details (telephone number/mobile number and e-mail address) for each listed company must also be submitted.</p>
---	--	--

MA



ENGR. ANA TRINIDAD F. RIVERA, MSc
 Chairperson, FDA-Bids and Awards Committee