

CENTRALIZED BIDS AND AWARDS COMMITTEE (CenBAC)

SUPPLEMENTAL BID BULLETIN NO.1

For ITB No. 2024-0041-CSL

Supply, Delivery, Installation, Commissioning, and Testing of Brand-New Spectrophotometers complete with all necessary parts and accessories for Common Services Laboratory (CSL) – Alabang and Cebu Testing and Quality Assurance Laboratories (ATQAL and CTQAL)

08 April 2024

This is in response to the issues/ concerns raised during the Pre-Bid Conference held on 4 April 2024 for the above-mentioned ITB, hereunder are the amendments, clarifications and recommendations.

Kindly refer to the comparative table given below for the previous (from) and revised (to) specifications:

A. INVITATION TO BIDS

FROM	TO
1. The <i>Food and Drug Administration</i> , through the Special Account in the General Fund (SAGF) intends to apply the sum of <i>Ten Million Seven Hundred Ninety Thousand Pesos Only (Php10,790,000.00)</i> being the Approved Budget for the Contract (ABC) to payments under the contract for <u>Supply, Delivery, Installation, Commissioning, And Testing of Brand-New Spectrophotometers complete with all necessary parts and accessories for Common Services Laboratory (CSL) – Alabang and Cebu Testing and Quality Assurance Laboratories (ATQAL and CTQAL)</u>	1. The <i>Food and Drug Administration</i> , through the Special Accounts in the General Fund (SAGF) – CY 2023 Continuing Appropriation (ConAp) intends to apply the sum of <i>Ten Million Seven Hundred Ninety Thousand Pesos Only (Php10,790,000.00)</i> being the Approved Budget for the Contract (ABC) to payments under the contract for <u>Supply, Delivery, Installation, Commissioning, And Testing of Brand-New Spectrophotometers complete with all necessary parts and accessories for Common Services Laboratory (CSL) – Alabang and Cebu Testing and Quality Assurance Laboratories (ATQAL and CTQAL)</u>
5. A complete set of Bidding Documents may be acquired by interested Bidders on 27 March 2024 to 16 April 2024 from the given address and website(s) below pursuant to the latest Guidelines issued by the GPPB, in the amount of Twenty – Five Thousand Pesos Only (Php25,000.00). The Procuring Entity shall	5. A complete set of Bidding Documents may be acquired by interested Bidders on 27 March 2024 to 16 April 2024 from the given address and website(s) below pursuant to the latest Guidelines issued by the GPPB, in the amount of Twenty – Five Thousand Pesos Only (Php25,000.00). The Procuring Entity shall



[Handwritten signature]

FROM	TO
<p>allow the bidder to present its proof of payment for the fees.</p>	<p>allow the bidder to present its proof of payment for the fees.</p> <p>Payment Options:</p> <ol style="list-style-type: none"> 1. Order of payment may be secured through the CenBAC Secretariat onsite. Payment shall be made at the FDA Main Office Cashier Section for prospective bidders from Metro Manila and Luzon, from 9:00am –12:00 noon. 2. For prospective bidder(s) from Visayas and Mindanao Region: <ol style="list-style-type: none"> a. Prospective bidder shall request/secure assessment slip from the CenBAC Secretariat through e-mail. Kindly indicate the ITB number and the number of item/s to be bid (e.g., 2 out of 4 items or all items) b. Upon issuance of assessment slip, payment shall be made at any Development Bank of the Philippines (DBP) using the below account details: <p style="margin-left: 40px;">Account name: FDA Academy Trust Account Account number: 00-0-00291-430-9</p> c. A scanned copy of proof of payment (validated deposit slip) shall be provided to the BAC Secretariat using the same email thread. This scanned copy of proof of payment will be endorsed to the FDA Cashier Section for payment verification
<p>7. Bids must be duly received by the CenBAC Secretariat through manual submission on or 16 April 2024, 08:30AM at the Food and Drug Administration CenBAC Secretariat Room, FDA Annex Bldg. Late bids shall not be accepted. (Note: Scanned copy of bids shall be provided and enclosed within the bid envelope. Files shall be saved in a USB flash drive and password protected)</p>	<p>7. Bids must be duly received by the CenBAC Secretariat through manual submission on or 17 April 2024, 08:30AM at the Food and Drug Administration CenBAC Secretariat Room, FDA Annex Bldg. Late bids shall not be accepted. (Note: Scanned copy of bids shall be provided and enclosed within the bid envelope. Files shall be saved in a USB flash drive and password protected)</p>
<p>9. Bid opening shall be on 16 April 2024 at 09:00 AM at Food and Drug Administration CenBAC Secretariat Room, FDA Annex Bldg. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>	<p>9. Bid opening shall be on 17 April 2024 at 09:00 AM at Food and Drug Administration CenBAC Secretariat Room, FDA Annex Bldg. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.</p>



Handwritten signature

Handwritten signature


B. SECTION II. INSTRUCTION TO BIDDERS

FROM	TO
2. Funding Information 2.2 The source of funding is: Special Allotment in the General Fund (SAGF) – CY 2023 Continuing Appropriation (ConAp)	2. Funding Information 2.2 The source of funding is: Special Accounts in the General Fund (SAGF) – CY 2023 Continuing Appropriation (ConAp)

C. SECTION VII TECHNICAL SPECIFICATION

ITEM NO.1 and 2: UV-Vis SPECTROPHOTOMETER

FROM	TO
A.2 Monochromator: Czerny-Turner Monochromator	A.2 Monochromator: Czerny-Turner Monochromator or equivalent/better <i>Note: Literatures / documents that are equivalent or better must be provided during bid submission.</i>
A.7 Stray Light	Less than or equal to 0.03% T for either 220 nm or 390 nm
C.1 Additional Spare Parts <ul style="list-style-type: none"> 1 pc additional/spare lamp (Note: Not applicable if light source offer is Xe Lamp) 1 pair of standard glass cuvettes with cover, 10mm pathlength 1 pair quartz cuvettes with cover, 10mm pathlength 1 set auto sipper unit or equivalent (if applicable) 1 set Multi Cell Sample Compartment with at least 6 sample placers including the reference standard 	C.1 Additional Spare Parts <ul style="list-style-type: none"> 1 pc additional/spare lamp (Note: Not applicable if light source offer is Xe Lamp) At least 1 pair of standard glass cuvettes with cover, 10mm pathlength At least 1 pair quartz cuvettes with cover, 10mm pathlength 1 set auto sipper unit or equivalent (if applicable) 1 set Multi Cell Sample Compartment with at least 6 sample placers including the reference standard
D.2 Provision of at least 6KVA online single-phase UPS. Preferably, the UPS should include visual and/or audible alarm capabilities. <i>Note: Brochure must be provided upon submission of bid</i>	D.2 Provision of at least 6KVA online single-phase UPS. Preferably, the UPS should include visual and/or audible alarm capabilities. <i>Note: Brochure and/or web link must be provided upon submission of bid</i>


ITEM NO. 3: ATOMIC ABSORPTION SPECTROPHOTOMETER WITH FLAME AND GRAPHITE FURNACE ATOMIZER

FROM	TO
ITEM NO. 3: ATOMIC ABSORPTION SPECTROPHOTOMETER WITH GRAPHITE FURNASE ATOMIZER	ITEM NO. 3: ATOMIC ABSORPTION SPECTROPHOTOMETER WITH FLAME AND GRAPHITE FURNACE ATOMIZER
A.1. Optics	
Optical System/Monochromator: Czerny-Turner or Echelle Type Monochromator	Optical System/Monochromator: Czerny-Turner or Littrow or Echelle Type Monochromator
Detector: Photomultiplier tube or equivalent	Detector: Photomultiplier tube or equivalent/ better <i>Note: Literatures / documents that are equivalent or better must be provided during bid submission.</i>
A.2. Flame Atomizer	
Burner type capable to: pre-mix (air-acetylene, nitrous oxide-acetylene)	Burner type capable to: pre-mix (air-acetylene and nitrous oxide-acetylene)
Safety Features: Flame monitoring Gas and/or water pressure monitoring Power failure protection	Safety Features: Flame monitoring Gas and/or water pressure monitoring or management Power failure protection
A.3 Graphite Furnace	
Safety Features: Furnace temperature monitoring Gas and/or water pressure monitoring	Safety Features: Furnace temperature monitoring Gas and/or water pressure monitoring or management
C.1. Recirculating chiller compatible to the main unit: <ul style="list-style-type: none"> • Temperature Range: Should include 5°C to 40°C • Cooling Capacity: At least 1000 watts (60Hz) • Temperature Stability: At least ± 0.2°C 	C.1. Recirculating chiller compatible to the main unit: <ul style="list-style-type: none"> • Temperature Range: Should include 5°C to 30°C • Cooling Capacity: At least 900 watts (60Hz) • Temperature Stability: Less than ± 0.2°C
C.3. • at least one (1) set Gas liquid separator	C.3. • at least one (1) set Gas liquid separator or equivalent
E.1 A minimum of one (1) year warranty on parts and services of GFAA System, Computer Set Parts, Printer, and UPS, commencing after successful installation by	E.1 A minimum of one (1) year warranty on parts and services of Atomic Absorption Spectrophotometer with Flame and Graphite Furnace Atomizer, Computer Set



[Handwritten signature]

[Handwritten signature]

FROM	TO
<p>Supplier and its acceptance by FDA. During the warranty period, Supplier must repair or replace any system component on site at no charge to FDA. Labor and travel must be included in the warranty program.</p>	<p>Parts, Printer, and UPS, commencing after successful installation by Supplier and its acceptance by FDA. During the warranty period, Supplier must repair or replace any system component on site at no charge to FDA. Labor and travel must be included in the warranty program.</p>
<p>J. SITE INSPECTION</p> <p>Site inspection for CTQAL is scheduled on 05 April 2024, from 09:00 AM to 4:00 PM. To ensure smooth operations and prevent overcrowding during the inspection, all interested suppliers are required to coordinate their visit to the laboratory with the Centralized Bids and Awards Committee (CenBAC) Secretariat via e-mail at bacsec@fda.gov.ph or directly through their phone number: (02) 8857-1900 local 8307 to secure an allocated time slot.</p>	<p>J. SITE INSPECTION</p> <p>Site inspection for ATQAL is scheduled on 05 and 08 April 2024, with two sessions each day at 09:00 AM and 01:30PM. To ensure smooth operations and prevent overcrowding during the inspection, all interested suppliers are required to coordinate their visit to the laboratory with the Centralized Bids and Awards Committee (CenBAC) Secretariat via e-mail at bacsec@fda.gov.ph or directly through their phone number: (02) 8857-1900 local 8307 to secure an allocated time slot.</p>


ENGR. ANA TRINIDAD F. RIVERA, MSc.
 Chairperson, CenBAC

